Features

Regulated Converter

- 1.8"x3.2"x1.2", encapsulated module
- 40W power from -40°C up to +65°C ambient
- Operating temp. up to +85°C with derating
- 4 kVac/1min reinforced isolation
- 2MOPP medical certified, B and BF compliant
- 5000m (medical/ITE) operating altitude
- Class B EMC filter built-in

Description

The ultra-compact encapsulated industrial + household + medical grade AC/DC converter series RACM40-K delivers 40 watts of output power from -40°C to +65°C with natural air convection only, and up to +85°C with derating or forced air cooling. With a clear focus on extended thermal performance for systems where space is limited, these 1.8" \times 3.2" compact modules are designed to gain highest overall efficiency levels over the full output load range from universal AC inputs. The RACM40-K has ANSI/AAMI/IEC 60601-1 medical safety and EN 60601-1-2 medical EMC certifications and offers 4kVac/1 min isolation, 2MOPP, and is designed to meet B and BF requirements. It is additionally certified (CB Report) to IEC/EN 62368-1; IEC61010 and IEC61558-1/-2-16 for industrial applications and IEC/EN 60335-1 for household appliances. The robust built-in class B EMC filter has sufficient margin to allow either Class II or Class I PELV with grounded output installations. The mechanically rugged construction with fully potted encapsulation, 1,6mm pins and additional threaded inserts gives the series enhanced stability against shock and vibrations.

Selection Guide					
Part Number	Input Voltage Range [VAC]	Output Voltage [VDC]	Output Current [mA]	Efficiency typ. ⁽¹⁾ [%]	Max. Output Power [W]
RACM40-05SK-T	80-264	5	6000	87	30
RACM40-12SK-T	80-264	12	3334	90	40
RACM40-15SK-T	80-264	15	2667	90	40
RACM40-24SK-T	80-264	24	1667	90	40
RACM40-48SK-T	80-264	48	833	90	40

Notes:

Note1: Efficiency is tested at +25°C with constant resistant mode at full load and 230VAC

Selection Guide	(on request MO	Q ≥1008pcs)		
Part Number	Input Voltage Range [VAC]	Output Voltage [VDC]	Output Current [mA]	Efficiency typ. ⁽¹⁾ [%]	Max. Output Power [W]
RACM40-18SK-T	80-264	18	2222	90	40
RACM40-36SK-T	80-264	36	1111	90	40

Model Numbering





RACM40-K

40 Watt
1.8" x 3.2"



Single Output

















CB Report (pending)



IEC/EN62368-1 (pending)
ANSI/AAMI ES60601-1 certified
CSA/CAN-C22.2 No. 60601-1:14 certified
IEC/EN60601-1 certified
IEC/EN60335-1 (pending)
IEC/EN61010-1 (pending)
EN62233 (pending)
IEC/EN61558-1 (pending)
IEC/EN61558-2-16 (pending)
EN55032/35 compliant
IEC/EN60601-1-2 compliant



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)


Parameter	Condition		Min.	Тур.	Max.	
Naminal Innut Valtage	60Hz		100VAC			
Nominal Input Voltage		50	OHz			240VAC
Operating Range (2)		47-0	63Hz	80VAC		264VAC
			OC	120VDC		370VDC
Input Current		115	5VAC			1000mA
input Guirent		230	OVAC			500mA
Inrush Current	cold	etart	115VAC			15A
illiusii Guireit	Colu	start	230VAC			30A
No load Power Consumption		230	OVAC		100mW	
	115VAC	RACM40	input power max. 0.5W	0.3W		
ErP Standby Mode Conformity		RACM40 input power max. 1.0W		0.7W		
(Maximum output power available for stated maximum input power)	0001/40	RACM40	input power max. 0.5W	0.27W		
maximum input power)	230VAC RACM40 input power max. 1.0W		0.65W			
Input Frequency Range				47Hz		63Hz
Minimum Load			0%			
Davier Frater		115	5VAC	0.6		
Power Factor		230	OVAC	0.5		
Start-up Time					160ms	
Rise Time					70ms	
Hold up Time	115VAC		5VAC	16ms		
Hold-up Time		230VAC		60ms		
Internal Operating Frequency		100% load a	at nominal Vin		100kHz	
Output Ripple and Noise (3)	20M	J- D\//	5Vout			80mVp-p
Output hippie and noise	20MHz BW others				1% of Vou	


Notes:

Note2: The products were submitted for safety files at AC-Input operation

Note3: Measurements are made with a $0.1\mu F$ MLCC & $10\mu F$ E-cap in parallel across output. (low ESR)

Efficiency vs. Load







Series


Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)


REGULATIONS				
Parameter	Cond	lition	Value	
Output Accuracy			±1.0% typ. / ±2.0% max.	
Line Regulation	low line to high line	5Vout others	$\pm 0.1\%$ typ. $\pm 0.05\%$ typ.	
Load Regulation (4)	100/ to 1000/ load	5, 12, 15Vout	0.7% typ.	
	10% to 100% load	24, 48Vout	0.5% typ.	
T : 10	25% load s	step change	3.0% max.	
Transient Response	recove	recovery time 500p	500μs max.	


Notes:


Note4: Operation below 10% load will not harm the converter, but specifications may not be met

Deviation vs. Load









PROTECTIONS		
Parameter	Туре	Value
Internal Input Fuse		T3.15A, slow blow type
Short Circuit Protection (SCP)	below 100mΩ	hiccup, auto recovery
Over Voltage Protection (OVP)		105% - 120% of nom. Vout, hiccup mode
Output reverse Voltage Protection	overrun rate of nominal output	107% - 145% of nom. Vout, hiccup mode
	continued on next page	



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

Parameter	Type/Condition		Value
Over Current Protection (OCP)			130% - 180% of nom. lout, hiccup mode
Thermal Shutdown	measured on TC point refer to	"Dimension Drawing (mm)"	+130°C typ.
Over Voltage Category (OVC)			OVCII
Class of Equipment			Class II
Isolation Voltage (safety certified) (5)	I/P to O/P	1 minute	4kVAC
Isolation Resistance	I/P to O/P	I/P to O/P, Viso= 500VDC	1G Ω min.
Isolation Capacitance	I/P to O/P	I/P to O/P, 100KHz/0.1V	100pF max.
Leakage Current			1.5mA max.
Insulation Grade			reinforced

Notes:

Note5: For repeat Hi-Pot testing, reduce the time and/or the test voltage


ENVIRONMENTAL				
Parameter	Condition		Value	
Operating Temperature Range	@ natural convection 0.1m/s	without derating	-40°C to +60/65°C	
Operating remperature hange	(refer to "Derating Graph")	with derating	-40°C to +85°C	
Max. Case Temperature			100°C	
Temperature Coefficient			±0.02%/K	
Thermal Impedance			6.3K/W	
Operating Altitude (6)	according to 62368-1/61010 and 60601-1		5000m	
Operating Humidity	non-condensing		20% - 95% RH max.	
Pollution Degree			PD2	
Vibration	according to MIL-STD-202G		10-500Hz, 2G 10min./1cycle, period 60min. along x,y,z axes	
MTBF	according to MIL-HDBK-217F, G.B.	+25°C	>1006 x 10 ³ hours	
IVIIDF	according to Mile-FIDBN-217F, G.B.	+40°C	>790 x 10 ³ hours	
Design Lifetime	230VAC/60Hz and full load +40	O°C	>98 x 10 ³ hours	


Notes:

Note6: Recognized by safety agency for safe operation up to 5000m. High altitude operation may impact the performance and lifetime. Please contact RECOM tech support for advice

Derating Graph

(@ Chamber and natural convection 0.1 m/s)





Notes:

Note7: Output power derating for Line-input of less than 90VAC (derate linearly from 100% at 90VAC to 80% at 80VAC)



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

SAFETY AND CERTIFICATIONS		
Certificate Type (Safety)	Report / File Number	Standard
Medical electrical equipment Part 1: General requirements for basic safety and essential performance	FF4400F D4004 4/A0/00 III	ANSI/AAMI ES60601-1:2005 + A2:2010/2012 CAN/CSA-C22.2 No. 60601-1:14, 3rd Edition
Medical electrical equipment Part 1: General requirements for basic safety and essential performance	E511305-D1001-1/A0/C0-UL	IEC60601-1:2005, 3rd Edition + AM1:2012 EN60601-1:2006 + A1:2013
Audio/Video, information and communication technology equipment - Safety requirements (CB Scheme)	pending	IEC62368-1:2014 2nd Edition
Audio/Video, information and communication technology equipment - Safety requirements (LVD)	pending	EN62368-1:2014 + A11:2017
Household and similar electrical appliances - Safety - Part 1: General requirements	pending	IEC60335-1:2010 5th Edition + C1:2016
Household and similar electrical appliances — Safety — Part 1: General requirements (LVD)	pending	EN60335-1:2012 + A14:2019
Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements (CB Scheme)	pending	IEC61010-1:2010+A1:2016, 3rd Edition
Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements	pending	EN61010-1:2010+A1:2019
Measurement methods for electromagnetic fields of household appliances and similar apparatus with regard to human exposure	pending	EN62233:2008
Safety of power transformers, power supplies, reactors $\&$ similar products for supply voltages up to 1100V (CB Scheme)	pending	IEC61558-1:2005 2nd Edition + A1:2009
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100V	pending	EN61558-1:2005 + A1:2009
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100 V Part 2: Particular requirements (CB Scheme)	pending	IEC61558-2-16:2009 1st Edition + A1:2013
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100 V Part 2: Particular requirements	pending	EN61558-2-16:2009 + A1:2013
RoHS2		RoHS 2011/65/EU + AM2015/863
EMC Compliance (Medical)	Condition	Standard / Criterion
Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests 4th Ed.	4789293779	EN60601-1-2:2015
ESD Electrostatic discharge immunity test	Air ±2, 4, 8, 15kV; Contact ±8kV	IEC61000-4-2:2008 , Criteria A EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	9V/m (710, 745, 780, 5240, 5500, 5785MHz) 10V/m (80-2700MHz) 27V/m (385MHz) 28V/m (450, 810, 870, 930, 1720, 1845, 1970, 2450MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Por:t L, N, L-N ±2kV	IEC/EN61000-4-4:2012, Criteria A
Surge Immunity	AC Port L-N: ±0.5, 1, 2kV L-PE, N-PE: ±0.5, 1, 2, 4kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 3Vrms (0.15-80MHz) 6Vrms (IMS Band)	IEC61000-4-6:2013, Criteria A EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	30A/m	IEC61000-4-8:2009, Criteria A EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions	Voltage Dips 30% Voltage Dips 100% (0.5P) Voltage Dips 100% (1.0P) Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria B
continued of	on next page	



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)


EMC Compliance (Industrial)	Condition	Standard / Criterion
Electromagnetic compatibility of multimedia equipment – Emission Requirements	L 00000010044DE	EN55032:2015
Electromagnetic compatibility of multimedia equipment – Immunity requirements	LCS200616044BE	EN55035:2017
ESD Electrostatic discharge immunity test	Air ±2, 4, 8kV; Contact ±2, 8kV	IEC61000-4-2:2008 , Criteria A EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	3V/m (4800-1000MHz, 1800, 2600, 3500, 5000MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Port: L, N, L-N ±1kV	IEC/EN61000-4-4:2012, Criteria B
Surge Immunity	AC Port: L-N: ±1kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 3Vrms (0.15-80MHz) 3Vrms (10-30MHz) 1Vrms (30-80MHz)	IEC61000-4-6:2013, Criteria A EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	1A/m	IEC61000-4-8:2009, Criteria A EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions	Voltage Dips 30% Voltage Dips 100% Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria C IEC/EN61004-11:2004, Criteria B IEC/EN61004-11:2004, Criteria C
EMC Compliance (Low voltage power supply)	Condition	Standard / Criterion
Low voltage power supplies, d.c. output Part 3: Electromagnetic compatibility (EMC)	LCS200616049BE	IEC/EN61204-3:2018
ESD Electrostatic discharge immunity test	Air ±2, 4, 8kV; Contact ±2, 8kV	IEC61000-4-2:2008, Criteria A EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	10V/m (80-1000MHz) 3V/m (1400-2000MHz) 1V/m (2000-2700MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Port: L, N, L-N ±2kV	IEC/EN61000-4-4:2012, Criteria B
Surge Immunity	AC Port: L-N: ±1kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 10Vrms (0.15-80MHz)	IEC61000-4-6:2013, Criteria A EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	30A/m	IEC61000-4-8:2009, Criteria A EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions	Voltage Dips 20, 30,60% Voltage Dips 100% (0.5P) Voltage Dips 100% (1.0P) Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria C IEC/EN61004-11:2004, Criteria B IEC/EN61004-11:2004, Criteria B IEC/EN61004-11:2004, Criteria C
Limits of Voltage Fluctuations & Flicker		EN61000-3-3:2013
Limitations on the amount of electromagnetic interference allowed from digital and electronic devices		FCC 47 CFR Part 15 Subpart B, Class B
Limitations on the amount of electromagnetic interference allowed from digital and electronic devices, industrial, scientific, and medical equipment		FCC 47 CFR Part 18

Parameter	Туре	Value
	PCB	FR4, (UL94 V-0)
Material	potting	PU, (UL94 V-0)
	baseplate	plastic, (UL94V-0)
Dimension (LxWxH)		83.23 x 46.43 x 30.40mm
Weight		185g typ.



Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)



PACKAGING INFORMATION			
Parameter	Туре	Value	
Packaging Dimension (LxWxH)	tray	365.0 x 210.0 x 56.0mm	
Packaging Quantity		12pcs	
Storage Temperature Range		-40°C to +90°C	
Storage Humidity	non-condensing	95% max.	

The product information and specifications may be subject to changes even without prior written notice. The product has been designed for various applications; its suitability lies in the responsibility of each customer. The products are not authorized for use in safety-critical applications without RECOM's explicit written consent. A safety-critical application is an application where a failure may reasonably be expected to endanger or cause loss of life, inflict bodily harm or damage property. The applicant shall indemnify and hold harmless RECOM, its affiliated companies and its representatives against any damage claims in connection with the unauthorized use of RECOM products in such safety-critical applications.